



# **UL DUTY STATION**

### **APPLICATION**

The LI-150B Duty Station provides an indication of incoming calls in work areas away from the nurse call master station.

Normal calls are indicated by a steady light and intermittent tone.

Emergency calls take precedence and are indicated by a flashing solid-state LED light and a pulsating tone.

A tone-off switch permits silencing of the tone signal for normal calls only.

### **FEATURES**

- UL® 1069 Listed
- Simple, reliable operation
- · Call-placed indicator
- Mounted on rugged, flame-retardant (UL® 94V-0) plastic panel

## **SPECIFICATIONS**

Dimensions: Height: 4.5" (114 mm)

Width: 2.75" (70 mm)
Depth: 1" (25 mm)

Projects: 0.5" (13 mm) from mounting surface

Construction: Reliable components mounted on a

flame-retardant (UL® 94V-0) plastic faceplate

Connections: Screw terminals

Housing: Single gang ring (IH-357) with squarebox

(IH-358), or exact equal.

### REQUIRED COMPONENTS

Tek-CARE® NC-200, NC-150 or NC-110 Nurse Call System IR-150 series or SF-100 series Patient Stations



LI-150B Duty Station

# ARCHITECTS' AND ENGINEERS' SPECIFICATIONS

Duty Stations shall be TekTone® LI-150B or approved equal as supplied by *LEE DAN®*. Both visual and audible indication of normal and emergency call shall be provided. A solid-state LED indicator shall provide a steady signal for normal calls and a flashing signal for emergency calls.

A **TONE OFF** switch shall be provided, which shall not be able to cancel an emergency signal. The station shall consist of reliable components mounted on a durable fireretardant (UL® 94V-0) plastic panel, and shall have screw terminals for electrical connections.

The LI-150B shall be UL® 1069 Listed.

800-231-1414 • 631-231-1414 155 Adams Avenue • Hauppauge, NY 11788-3699 Fax 631-231-1498 • www.*LEEDAN*.com INTERCOMS • MAILBOXES • DIRECTORIES
NURSE-CALL • ANNUNCIATORS • WIRE & CABLE